USER MANUAL

S12









HEARTWAY MEDICAL PRODUCTS CO., LTD.

NO. 6, ROAD 25, TAICHUNG INDUSTRIAL PARK, TAICHUNG. TAIWAN R.O.C.408

PUBLISHED: SEP.6TH, 2012 VERSION: 1


PART NO.: 70030175

CONTENTS

1.	SAFETY INSTRUCTION
2.	ELECTOMAGNETIC INTERFERENCE(EMI)7
3.	SPECIFICATIONS
4.	PRODUCT DIMENSION10
5.	COMFORT ADJUSTMENT
6.	OPERATION OF CONTROL PANEL
7.	CHARGING INSTRUCTION
8.	BATTERY INSTRUCTION & MAINTENANCE
9.	SCOOTER MAINTENANCE & REPAIR
10.	TROUBLESHOOTING & FAULT REPAIR
11.	CIRCUIT DIAGRAM
12.	. HEARTWAY'S PATENT-SUSPENSION SYSTEM
13.	BOM LIST DRAWING
14.	WARRANTY DECLARATION40

SAFETY INSTRUCTION

♦ General



Do not attempt to climb curbs greater than	Do not leave your hands and legs off the scooter
limitation show on Technical Specification	when driving.
Do not ride your scooter during snow in order to	Do not allow unsupervised children to play near
avoid accident on slippery road.	this equipment while the batteries are charging.

- Warning Don't operate your scooter for the first time without completely reading and understanding this user manual.
- 1. Don't operate scooter on public streets and roadways. Be aware that it may be difficult for traffic to see you when you are seated on the scooter. Obey all local pedestrian traffic rules. Wait until your path is clear of traffic, and then proceed with extreme cautions.
- 2. To prevent injury to yourself or others, always ensure that the power is switched off when getting on or off of the scooter.
- 3. Always check that the drive wheels are engaged (drive mode) before driving. Do not switch off the power when the scooter is still moving forward. This will bring the chair to an extremely abrupt stop.
- 4. Do not use this product or any available optional equipment without first completely reading and understanding these instructions. If you are unable to understand the warnings, cautions or instructions, contact a healthcare professional, the dealers or technical supports before attempting to use this equipment, otherwise, injury or damage may occur.

- 5. There are certain situations, including some medical conditions, where the scooter user will need to practice operating the scooter in the presence of a trained attendant. A trained attendant can be defined as a family member or care professional especially trained in assisting a scooter user in various daily living activities. Consult with your physician if you are taking any medication that may affect your ability to operate your scooter safely.
- 6. Do not attempt to lift or move a power scooter by any of its removable parts including the armrests, seats or shrouds. Personal injury and damage to the power chair may result.
- 7. Never try to use your scooter beyond its limitations as described in this manual.
- 8. Please do not sit on your scooter while it is in a moving vehicle.
- 9. Keep your hands away from the wheels (tires) while driving scooters. Be aware that loose fitting clothing can become caught in the drive tires.
- 10. Consult your physician if you are taking prescribed medication or if you have any certain physical limitations. Some medications and limitations may impair your ability to operate scooters in a safe manner.
- 11. Be aware when the drive mode is unlocked or locked.
- 12. Don't remove anti-tipper if there is any-tipper equipped with the scooter.
- 13. Contact with tools can cause electrical shock and do not connect an extension cord to the AC/DC converter or the battery charger.
- 14. Do not attempt to lift or move your scooter by any of its removal parts, such as the armrests, seats, or shroud.
- 15. When climbing an incline, don't drive at an angle up the face of the incline. Drive your scooter straight up the incline. This greatly reduces the possibility of a tip or a fall.
- 16. Don't climb a slope steeper than the scooter's limitation.
- 17. Don't attempt to have your scooter proceed backward down any step, curb or other obstacle. This may cause the scooter to fall or tip.
- 18. Always reduce your speed and maintain a stable center of gravity when cornering sharply. Don't corner sharply when driving scooters at higher speeds.
- 19. Operating in rain, snow, salt, mist conditions and on icy or slippery surfaces may have an adverse affect on the electrical system.
- 20. Never sit on your scooter when it is being used in connection with any type of lift or elevation product. Your scooter is not designed with such use in mind and any damage or injury incurred from such use is not the responsibility of Heartway.

♦ Modifications

Heartway Medical Product has designed and engineered power scooters to provide maximum utility. However, under no circumstances should you modify, add, remove, or disable any part or function of your power scooter. Personal injury and damage to the power chair may result.

1. Do not modify your power scooter in any way not authorized by Heartway. Do not use accessories if they have not been tested or approved for Heartway products.

2. Get to know the feel of your power scooter and its capabilities. Heartway recommends that you perform a safety check before each use to make sure your scooter operates safely.

• Inspections prior to using your power scooter:

- 1. If equipped with pneumatic tires, please check for proper tire inflations.
- 2. Please check all electrical connections and make sure they are tight and not corroded.
- 3. Please check all harness connections and make sure they are secured properly.
- 4. Please check the brakes.

♦ Weight limitation.

- 1. Please refer to the specifications table for weight capacity information. Power scooter is rated for a maximum weight capacity.
- 2. Stay within the specified weight capacity for your scooter. Exceeding the weight capacity voids your warranty. Heartway will not be held responsible for injuries or property damage resulting from failure to observe weight limitations.
- 3. Don't carry passengers on scooters. Carrying passengers on scooter may affect the center of gravity, resulting in a tip or a fall.

♦ Tire inflation

- 1. If your scooter is equipped with pneumatic tires, it is necessary to check the air pressure at least one time a week.
- 2. Proper inflation pressures will prolong the life your tires and ensure the smooth operation while riding.
- 3. Do not under-inflate or over-inflate your tires. It is critically important that 30-25 psi (2-2.4bar) tire pressure be maintained in pneumatic tires at all times.
- 4. Inflating your tires from an unregulated air source could over-inflate them, resulting in a burs tire.

♦ Temperature

- 1. Some of the parts of the power scooter are susceptible to change in temperature. The controller can only operate in temperature that ranges between -25° C ~ 50° C.
- 2. At extreme low temperatures, the batteries may freeze, and your power scooter may not be able to operate. In extreme high temperatures, it may operate at slower speeds due to a safety feature of the controller that prevents damage to the motors and other electrical components.

ELECTROMAGNETIC INTERFERENCE (EMI)

The rapid development of electronics, especially in the area of communications, has saturated our environment with electromagnetic (EM) radio waves that are emitted by television, radio and communication signals. These EM wave are invisible and their strength increases as one approach the source. All electrical conductors act as antennas to the EM signals and, to varying degrees, all power wheelchairs and scooters are susceptible to electromagnetic interference (EMI). The interference could result in abnormal, unintentional movement and/or erratic control of the vehicle. The United States Food and drug Administration (FDA) suggests that the following statement be incorporated to the user's manual for all power scooters like the S12, S12T, S12S, S12X. Power scooters may as susceptible to electromagnetic interference (EMI), which is interfering electromagnetic energy emitted from sources such as radio stations, TV stations, amateur radio (HAN) transmitter, two-way radios, cellular phones and alarm systems of shops. The interference (from radio wave sources) can cause the power scooter to release its brakes, move by itself or move in unintended directions. It can also permanently damage the powered scooter's control system. The intensity of the EM energy can be measured in volts per meter (V/m). Each powered scooter can resist EMI up to a certain intensity. This is called "immunity level". The higher the immunity level the greater the protection. At this time, current technology is capable of providing at least 20 V/m of immunity level, which would provide useful protection against common sources of radiated EMI.

Following the warnings listed below should reduce the chance of unintended brake release or powered scooter movement that could result in serious injury:

- 1. Do not turn on hand-held personal communication devices such as citizens band (CB) radios and cellular phones while the powered scooter is turned on.
- 2. Be aware of nearby transmitters such as radio or TV stations and try to avoid coming close to them.
- 3. If unintended movement or brake release occurs, turn the powered scooter off as soon as it is safe.
- 4. Be aware that adding accessories or components, or modifying the powered scooter, may make it more susceptible to interference from radio wave sources (Note: It is difficult to evaluate the effect on the overall immunity of the powered scooter).

5. Report all incidents of unintended movement or brake release to the powered scooter manufacturer, and note whether there is a radio wave source nearby.

TURN OFF YOUR POWERED SCOOTER AS SOON AS POSSIBLE WHEN EXPERIENCING THE FOLLOWING:

- Unintentional scooter movements
- > Unintended or uncontrollable direction.
- Unexpected brake release


The FDA has written to the manufacturers of power scooters asking them to test new products to be sure they provide a reasonable degree of immunity against EMI. The FDA requires that a powered wheelchair should have an immunity level at least 20 V/m, which provides a reasonable degree of protection against more common sources of EMI. The higher the immunity level the greater the protection. Your powered scooter has an immunity level of 20 V/m which should protect against common sources of EMI. Warning: The scooter itself can disturb the performance of the electromagnetic fields such as emitted by alarm systems of shops.

S12/S12S/S12X - TECHNICAL SPECIFICATIONS


S12S VITA	S12 VITA	S12X VITA	V	12T ITA
MODEL	S12	\$12T	S12S	\$12X
		COMICS OF WINSHOP		
WEIGHT CAPACITY SEAT: TYPE/SIZE	160kgs(350 20" A2	160kgs(350 20" A2	160kgs(350 20" A2	160kgs(350 20" A2
DRIVE WHEEL	330mmx100m	330mmx100m	330mmx100m	20 A2 380mmx160m
DRIVE WHEEL	m(13"x5"")	m(13"x5")	m(13"x5"")	m(15"x6.3"")
FRONT CASTER (WHEEL)	330mmv100m	330mmx100m	330mmx100m	330mmx120m
TRONT CASTER (WILLE)	m(13"x4")	m(13"x4")	m(13"x4")	m(13"x4.8")
(ANTI-TIPPER)	None	None	None	None
MAX SPEED	8MPH	8MPH	11MPH	11MPH
	(13KM/H)	(13KM/H)	(18KP/H)	(18KP/H)
BATTERY SPECIFICATIO		80Ah x 2pcs	80Ah x 2pcs	80Ah x 2pcs
BATTERY RANGE	30km/50km	35km/45km	40km/50km	25km/45km
CHARGER TYPE	8Amp,Off	8Amp,Off	8Amp,Off	8Amp,Off
	Board120/240	Board120/240	Board120/240	Board120/240
	Volt, 50/60Hz	Volt, 50/60Hz	Volt, 50/60Hz	Volt, 50/60Hz
CONTROLLER TYPE	S-DRIVE	S-DRIVE	S-DRIVE	S-DRIVE
	120Amp	120Amp	140Amp	200Amp
MOTOR TYPE	4-Pole 700W	4-Pole 700W	4-Pole 700W	4-Pole 900W
WEIGHT: W/ BATTERY	135kgs(300	133kgs(293	140kgs(308	150kgs(330
WEIGHT: W/O BATTERY	95kgs(209 lbs)	93kgs(205 lbs)	100kgs(220 lbs)	110kgs(242 lbs
TURNING RADIUS	1280mm	1150mm	1280mm	1350mm
SUSPENSION	FULL	FULL	FULL	FULL
LENGTH	1420mm	1390mm	1420mm	1620mm
WIDE	700mm	700mm	700mm	790mm
HEIGHT	1360mm	1360mm	1360mm	1360mm
SEAT WIDTH	510mm	510mm	510mm	510mm
SEAT HEIGHT	460mm	460mm	460mm	460mm
SEAT DEPTH	480mm	480mm	480mm	480mm
BACK HEIGHT	750mm	750mm	750mm	750mm
WHEEL BASE	1025mm	980mm	1025mm	1075mm
GROUND CLEARANCE	80mm	80mm	80mm	100mm
LEG ROOM	320mm	340mm	320mm	320mm
				R




S12/S12S - DIMENSION



S12T - DIMENSION




S12X - DIMENSION



COMFORT ADJUSTMENT

Adjustments for Seating Comfort:




- Turn the swivel lever downwards to rotate the seat.
- Push the front lever upwards to move the seat forward and backward.

Driving and braking



- For driving, please pull the right throttle lever (forward) or the left throttle lever (backward).
- Please release the throttle lever to stop the scooter. In the center position the magnetic motor brake system works also as a parking brake. Optional there is a hand brake system available.
- For safety, when rolling at free-wheel mode down slope, the automatic brake will take action if the transferring speed is more than 30% of the scooter's maximum speed.



- Please be noted that the scooter will be at free-wheel mode, when the motor is disengaged.
- To use the parking brake, you must move and lock the lever into the engaged position!

Tiller Positioning

Press down the lever and move it to your preferred position.



The adjustment of rear suspension system:



The softness of the rear suspension system is adjustable. There are altogether 5-level of adjustments. It is applicable with around 120kg and each level can bear up to 13.5 kg.



OPERATION OF CONTROL PANEL

LCD (Liquid Crystal Display) Power Scooter Control Panel, TN Type

Model	LCD(Liquid Crystal Display)
Functions	1. Power Indicator : Battery remaining capacity and charging indicator (6 squares + Battery Icon)
	2. Clock : Hour / Minute / Second display and setting.
	3. Speed Sensor : 7 Segment display (2.5 digits +1 decimal) + "km/h / mph" symbol
	4. High/Low/ Turn Speed : Indicated as "H" and "L" symbols
	5. Odometer : ODO(99999km max) \ TRIP(99.9 max)
	6. Headlight : "Power-saving" mode, Blue LED
	7. Back-up Lamps : "Brake / Reverse" modes, Orange LED
	8. Right-Indicator : Flash mode, Green LED
l	9. Left-Indicator : Flash mode, Green LED
1	10. Parking Lamp : Including "Reverse Mode", left- indicator and right-indicator flashing
	simultaneously, Red LED
	11. Malfunction Code: 7 Segment display (1digit) + Warning symbol + Red LED
	12. Power-on Scan : All LED turn on
	13. TEMP Gauge : ℃ and ℉ modes
	14. Reverse Light : "Reverse" symbol flashing
Buttons	Eft-Indicator control
	High/low speed switch Parking

Buttons	: Horn
	: Headlight : MODE : SET : Back-up Light
LED Indicators	Right & Left indicator (Green) , Parking light (Red), Warning light (Red), Back-up lamps (Yellow),
	Headlight (Blue)
LCD Backlight	LED (White)
Connector	CON1: 20PIN AMP

Usage Condition

ITEM	SPECIFICATION
Voltage	DC24 V
Operation Voltage	DC 16 ~32 V
Storage TEMP.	$-40^{\circ}\text{C} \sim 65^{\circ}\text{C}$
Operation TEMP.	-25°C ~ 50°C
Meter Angle at	30° of elevation while scooter assembly (LCD orientate to six o'clock)
Handle Cover	

2 Seneral Characteristic Performance Test (20 ± 5℃)

2.1 • Headware Circuit:

ITEM	SPECIFICATION	RESULT
Lowest Operation VOLT	16 V max	V
Consuming Current $(V_B = 24.0V)$	Dynamic: 200 mA max Backlight and LED light status Static: 5 mA max Key OFF status	mA mA

1 \cdot Speed Sensor and Display

ITEM	SPECIFICATION		
Operation Features	Speed detection by speed sensor from transaxle with conversion at 1800rpm equal to		
	60km/h.		
Tolerance	5~15% (±2%)		
Digital Range	0.0 ~ 99		
Display Switch Button	Initial setting at km/h, switch to MPH by MODE and SET buttons		

2 • High / Low / Turn Speed

ITEM	SPECIFICATION	
Operation Features	(1) Switch High / Low speed by pressing button once. (TRN as control signals) Press	
	one time: High-speed <<>> Low-speed (with memory storage).	
	(2) Take exterior turn-switch as determinant signal (TRN as control signals).	
Symbols on LCD	" H" symbol means "High Speed"	
	" L" symbol means "Low Speed" :	
	Lspeed	
	" L" symbol flashing means "Turn Speed":	
	Lspeed	
Flicker Frequency	1 sec	

• Power Indication

ITEM	SPECIFICATION		
	Remaining Capacity (%)	Scale Bar	
	$ \begin{array}{c} 100 \\ (6) \\ \end{array} > 25.42 $		
	(5) ≤ 25.42		
Battery Remaining	$ \begin{array}{c} 70 \\ (4) \end{array} \leq \underline{25.12} \end{array} $		
Capacity	55 (3) ≤ 24.78		
	$ \begin{array}{c} 40 \\ (2) \end{array} \leq \underline{24.42} \end{array} $		
	$\begin{array}{l} 30\\(1) \end{array} \leq \underline{23.88} \end{array}$	F E E and Flashing	
	20 Low-power Warning	Warning LED Flashing	
Flicker Frequency	2 sec.		
Operation Characters	 (1) Scale status only decrease, won't increase. (2) When the remaining capacity was less than 30%, warning sound ("Be-Be" two short sounds) act at 5 seconds intervals. While (a) Key Off (b) Charging Mode (c) Sleep Mode, warning sound released. 		

ITEM	SPECIFICATION			
	Remaining Capacity (%)	Voltage (V)	Scale Bar	
	40 (2)	< 25.44		
	55 (3)	> 25.44		
Charge Indication	70 (4)	> 26.18		
	80 (5)	> 26.92		
	90 (6)	> 28.5		
	100 (7)			
Increase Frequency 0.5 sec.				
Operation Character		(CH3) of charg	n't increase. ger as determinant signal, enter 「Charging Mode」 when CH ON" or "KEY OFF".	13
Remarks	Above scale bar status only for reference, must take the indicator of charger as the precise diagnosis.		se	

4 Clock Meter

ITEM	SPECIFICATION
Tolerance (per day)	$\pm 2 \sec$
Initial Setting Value	『Hour:Min』 mode:『AM 12:00』
『 Hour : Min 』	Display range : AM12:00 ~ PM11:59
Setting	
(12-Hour format)	
	When 『Hour』 is between 1 and 9 o'clock, displayed at 1~9.

5 • Odometer

ITEM	SPECIFICATION			
Operation Features	Odometer detected by the signal of Opto Coupler then converts into distance.			
Display Switch Button	 └ km/h 」 means the odometer displayed as kilometer. └ mph 」 means the odometer displayed as mile. 			
Accumulative Display [ODO]	 (1) Display Range:00000~99999 ODO ODO			
TRIP Counter	 (1) Display Range : 00.0~99.9 TRIP 88.8 mile (2) When over 99.9km, display stop counting (won't restart from "00.0"). 			
Operation status(1) Odometer indication display on ODO mode when Power On, then switch to after 5 seconds. (2) TRIP can be reset to "00.0".				

6 • Headlight Control

ITEM	SPECIFICATION			
Operation Feature	 Take exterior headlight switch as determinant signal. (1) Switch on/off the head light by pressing button once, then LED will turn on/off simultaneously. (2) LCD backlights turn on / turn off with head light. 			
Power Saving Mode	When motor stop, the modulation down to 30% (Headlight) When motor act, 100% output power (Headlight)			
Usage Condition	While (a) KEY OFF (b) Power-Saving mode (c) Sleep mode, all functions closed.			
Determinant Condition(1) 2.2V>WIP>2.8V (100% Full-power) (2) 2.2V <wip>2.8V (100% Full-power) (3) Full / Half power switch at real time. (4) The determination of "Reversing Mode" need to consider the motor direction and setting.</br></wip>				
Remarks	(1) Loop Load: 24V/50W max(2) With "short circuit" and "overload" protection			

7 Sack-up Lamp control

ITEM	SPECIFICATION		
Operation Feature	 Take exterior back-up lamp switch as determinant signal. (1) Switch on/off the head light by pressing button once, then LED will turn on/off simultaneously. (2) LCD backlights turn on / turn off with head light. 		
(Control Mode) Brake-lamp Mode Reversing-lamp Mode	When motor changes from act (go forward) to stop, the lamp reinstated after flashing for 3 sec. Determine as "Reversing Mode", back-up lamp keep flashing. Reverse warning sound can be set by panel (Turn on / Turn off)		
Usage Condition	While (a) KEY OFF (b) Charging Mode (c) Sleep Mode, all functions closed. * Brake-lamp & Reversing-lamp Mode won't be limited by Back-up lamp switch on or off.		
Flicker Frequency	1 sec.		

ITEM	SPECIFICATION		
Determinant Condition	 (1) 2.2V>WIP>2.8V (50% Half-power) (2) 2.2V<wip>2.8V (100% Full-power)</wip> (3) Full / Half power switch at real time. (4) The determination of "Reversing Mode" need to consider the motor direction and panel setting. 		
Remarks	(1) Loop Load : 24V/50W max(2) With "short circuit" and "overload" protection		

$8 \cdot 9 \cdot 10$ Indicators and Parking Lamp Control

ITEM	SPECIFICATION			
Operation Feature	Take exterior left-right indicators and parking-lamps switch as the determinant signal.			
Control Mode (Left-direction lamp) (Right-direction lamp)	Press button once, the right-indicator and rescaled again to turn off, left-indicator and again to turn off left-indicator. Press button once, the right-indicator and rescaled again to turn off, left-indicator and rescaled again to turn off, left-indicator and rescaled again to turn off left-indicator. Press button once, the right-indicator and rescaled again to turn off left-indicator and rescaled again to turn off left-indicator. Press button once, warning sound act. Press again to turn off left-indicator. Press button once, turn on, right-left indicators and rescaled again to turn off left-indicator.			
(Parking lamp)				
Usage Condition	While (a) KEY OFF (b) Charging Mode (c) Sleep Mode, all functions closed.			
Flicker Frequency	1 sec.			
Warning Sound Frequency	One short "Bi" sound per second			
Remarks	 (1) Load circuit for left-direction light: 24V/50W max (2) Load circuit for right-direction light: 24V/50W max (3) With "short circuit" and "overload" protection 			

11 Nalfunction Message

ITEM	SPECIFI	SPECIFICATION			
Operation Feature		Take the connector pin (KEY) of controller as determinant signal, then converts it into digital code.			
		When the controller send out an error message, red LED flashing with controller signal at same time, the "Error message code" will show on LCD.			
Usage Condition	ERŘU	km/h			
Flicker Frequency	1 sec.				
Controller message (Flicker)	Message code	ERROR symbol	(Flicker)	<u>Status</u>	
1				Battery needs charge soon.	
2	2	On		Low-voltage, needs charge now	
3	3	On	Flashing, opposite to controller	Over-voltage	
4	4	On On		Over-current	
5	5			Park Brake lost or faulted	
6	6	On	message.	Accelerator not align center	
7	7	On		Accelerator broken or faulted	
8	8	On		Motor broken or faulted	
9	9	On		Others	

1 2 • Power on Self-Test

ITEM	SPECIFICATION		
Initial Status	When scooter power on, the control panel will go through a self-test routine; the backlight and all LCD segments will be tuned on for 3 seconds, then switch automatically to the general operation mode (ODO).		

1 3 • Temperature Sensor

ITEM	SPECIFICATION		
Operation Feature	Temperature detected by temperature sensor (NTC) from transformation with signal.		
Tolerance	± 2°C		
Display Range	-20°C ~50°C -4°F ~122°F TEMP		
Display Switch Button	When display °C, degree stand for Celsius thermometer When display °F, degree stand for Fahrenheit thermometer		

14 · Reverse Indicator

ITEM	SPECIFICATION		
Operation Feature	Take exterior forward / backward switch as determinant signal.		
Power Saving Mode	When switch direct to "forward", no symbol on LCD. When switch direct to "backward", symbol flashing on LCD.		
Flicker Frequency	1 sec.		

15 · Buttons


TEM SPECIFICATION				
Button	"MODE" switch Function set			
General Display Mode (TRIP)	Press SET for 3 seconds to reset TRIP at "00.0".			
Setting Mode	 Press MODE and SET simultaneously for more than 2 seconds. to enter "Setting Mode", then "Hour : MIN a start flashing. (1) When "Hour a flashing: Press SET to increase of number, then press MODE to enter "Setting Mode" of "MIN a. (2) When "MIN a flashing: Press SET to increase of number, then press MODE to enter "Setting Mode" of "km/h & mph a. (3) When "km/h a or "mph a flashing Press SET to choose "km/h" or "mph" type, then press MODE to enter "Setting Mode" of "Setting Mode" of "°C / °F a (4) When "°C a or "°F a flashing Press SET to choose °C or °F . 			
Escape from Setting Mode	Under setting mode, if below situations happened, will auto save the last setting value then escape to general operation mode. (1) No any operation of ADJ button for 20 sec. (2) Press MODE and SET at same time for more than 2 sec.			
 (1) 『Hour: Min』, 『km/h』 or 『mph』, 『°C』 or 『°F』 offer Cyclical Switch fu (2) When adjusting 『Hour: Min』, press SET to increase number, if press S more than 2 seconds, the number will increase continuously until button released value with Cyclical Switch function (only 2 seconds from 0 to 9). * If 『Hour』 less than 10, the denary "0" doesn't display. ° 				
Remarks	Button tones: one short "Bi" sound			

CHARGING INTRUCTIONS

Battery Charger Instruction



1. APPEARANCE



Dewer Cord
Output Plug to Battery
Indicator:

Green Flash: Power On
Orange Flash: Pre Charge
Orange: Charging
Green & Orange Flash: Charged 80%
Green: Full Charged
Red Flash : Defected

2. SPECIFICATION

Item	BATTERY CHARGER (SWITCHING MODE)		
Model	4C24080A		
Output Current(DC)	8A±5%		
Charging Voltage(DC)	28.8V		
Floating Voltage(DC)	27.6V		
Input Current (AC)	3.8A max.		
Input Voltage(AC)	100 ~ 240 V 50/60Hz		
Efficiency	AC-DC 85% min		
Operating Temperature	$0^{\circ}C \sim 40^{\circ}C$		
Switching Method	SWITCHING MODE		
Charging Method	Constant current two stage constant voltage		
Battery Application	24V Lead Acid Rechargeable Battery (26Ahr ~ 75Ahr)		
	1.Short Circuit Protection		
Output Detection	2.Reverse Power Protection		
Output Detection	3. Overheat Protection		
	4. Charging Plug Protection		
Operating Humidity	20% ~ 85 %		
Measure L 185mm×W 130mm×H 195mm			
Weight	1.7K g		
Color	Blue		

3. OPERATING INSTRUCTION

- (1)Make sure the battery charger output voltage is the same as the connecting battery.
- (2)Plug in the power cord. LED indicates green flash when AC power on.
- (3)Connect the battery charger to the battery.
- (4)Start charging; please refer to 4. LED INDICATION

4. LED INDICATION

- (1)Green Flash : Power on
- (2)Orange : Charging
- (3)Orange Flash : Pre charge
- (4)Green & Orange Flash : Charged 80% $\,\circ\,$
- (5)Green : Full charged(Floating charge) •
- (6)Red Flash : Defect

5. TROUBLE SHOOTING

- (1) If green indicator is off:
 - Check AC input. If it works functionally, the battery charger may be defective.
- (2) If green indicator keeps flashing and cannot turn to charging indication :
 - Check if the battery connector is connected successfully.
 - Check if there is any short circuit on the output connection.
 - The battery charger may be defective if the battery connection works functionally.

(3) If red indicator keeps flashing :

- Check if the battery connection is reversed.
- Check if there is any short circuit on the output connection.
- Check if the environment temperature is too low $(0^{\circ}C)$
- The battery charger may be defective if the red indicator still keeps flashing.
- (4) Charging indicator (orange) cannot turn to green :
 - The battery might be defective, please stop charging and have the battery be repaired.
- (5) If the charging indicator (orange) turns to green (fully charged) immediately :
 - The battery may be in well-charged condition.
 - The battery may be defective if the battery is not fully charged.

6. CAUTION

- (1) Before using the battery charger, read all instructions and cautionary markings.
- (2) Use the battery charger in a well-ventilated area
- (3) To avoid the risk of injury, charge only lead-acid or gel cell type rechargeable batteries.
- (4) Please turn off the power after charging



Important!

- Always charge your batteries in well ventilated areas.
- The charger is intended for indoor use only. Please protect it from the moisture.
- For maximum performance, it is recommended that you replace both batteries at the same time if the batteries are weak.
- If the scooter will not be used for a long period of time, arrange to have the batteries recharge at least once every month to avoid deterioration of the batteries.

BATTERY INSTRUCTION & MAINTENANCE

- Read through the charger operating instruction before using it.
- Make sure you charge the battery every time after you use the power chair or scooter.
- Charge the battery at least 24 hours a week if the power chair or scooter has not been used. (This is to make sure that the electrolyte is always at the top level)
- If the battery cannot be charged (Orange light cannot turn to Green) or if the Orange light turns to Green immediately, please check it with the technicians. The battery may be defective.
- The voltage difference between the two batteries on a power unit cannot be more than 0.5 V; the battery case should be inspected for cleanliness and evidence of damage.
- If the charger indicates red light, please kindly check if the charger is defected or if the cable wiring connection is poor.
- Please keep the battery Θ and \oplus connectors clean otherwise the charging condition will be poor.

SCOOTER MAINTENANCE & REPAIR

Your power scooter is designed for minimal maintenance. However, like any motorized vehicle it requires routine maintenance. To keep your <u>Vita Scooter</u> for years of trouble-free operation, we recommend you follow the following maintenance checks as scheduled.

DAILY CHECKS

- 1. Visual check on the conditions of tires.
- 2. Inspect the battery condition meter on the controller to determine if batteries need to be charged.

WEEKLY CHECKS

1. Your power scooter comes with standard pneumatic tires. If your power scooter comes with optional air tires, make sure to maintain the pressure of the tires between 30-35 psi.

MONTHLY CHECKS

1. Visually inspect the controller harnesses. Make sure that they are not frayed, cut or have any exposed wires.

SEMI-ANNUAL CHECKS

1. Check the motor brushes. We recommended that your authorized dealer inspect the brushes every six months or sooner if your power scooter is not operating smoothly. If inspection determines excessive wear on the brushes, they must be replaced or motor damage will result.

Warning! Failure to maintain the brushes could void the power scooter warranty.

To inspect or replace the motor brushes:

- 1. Unscrew the motor brush caps.
- 2. Remove the brushes.
- 3. Inspect the brushes for wear.
- 4. Replace the brushes if necessary.



Inspect the state of the battery terminals every six months. Make sure that they are not corroded and the connections are tight. Periodically apply a thin film of petroleum jelly on the surface of terminals to guard against corrosion.

CHECKS:

- Make sure to keep the controller clean while protecting it from rain or water. Never hose off your power scooter or place it in direct contact with water.
- Keep wheels free from lint, hair, sand and carpet fibers.
- Visually inspect the tire tread. If less than $1 \text{mm} (1/32^{\circ})$, please have your tires replaced by your local dealer.
- All upholstery can be washed with warm water and mild soap. Occasionally check the seat and back for sagging, cuts and tears. Replace if necessary. Do not store your scooter in damp or humid conditions as this will lead to mildew and rapid deterioration of the upholstery parts.
- All moving mechanism will benefit from simple lubrication and inspection. Lubricate using petroleum jelly or light oil. Do not use too much oil, otherwise small drips could stain and damage carpets and furnishings etc. Always perform a general inspection of the tightness of all nuts and bolts.

TROUBLESHOOTING & FAULT REPAIR

<u>RHINO controller</u>: Your scooter is fitted with a <u>Rhino controller</u>, which continuously monitors the operating conditions of your scooter. If it detects a problem it will indicate with error message by flashing light on the power ON/ OFF light. You must count the number of the flash, and see the list to check what kind of error has happened according to the number)

Number of Flashes	Fault	Impact on Scooter	Notes
1	Battery needs recharging	Will drive	Battery charge is running low. Recharge the batteries as soon as possible.
2	Battery voltage too low	Drive inhibited	Battery charge is empty. Recharge the batteries. If the scooter is left off for a few minutes, battery charge may recover sufficiently to allow driving for a short period of time.
3	Battery voltage too high	Drive inhibited	Battery charge is too high. If a charger is plugged in, unplug it or turn the Charge/Run switch to Run. Scooters powered by RHINO will charge the batteries when traveling down slopes or decelerating. Excessive charging in this manner may cause this fault. Turn the scooter power off and then back on again.
4	Current limit time out	Drive inhibited	The scooter has drawn too much current for too long, possibly because the motor has been over worked, jammed or stalled. Turn the scooter power off, leave for a few minutes, and then turn the power back on again. The controller has detected a shorted motor. Check the loom for shorts and check the motor. Contact your service agent.
5	Brake fault	Drive inhibited	Check that the park brake release lever is in the engaged position. The park brake coil or wiring is faulty. Check the park brake and wiring for open or short circuits. Contact your service agent.

6	Out of Neutral at Power Up	Drive inhibited	Throttle is not in neutral position when tuning switch key on. Return throttle to neutral, turn power off and back on again. Throttle may need to be re-calibrated Check throttle wiring.
7	Speed Pot Error	Drive inhibited	The throttle or its wiring is faulty. Check for open or short circuits. Throttle may not be correctly set up. Contact your service agent.
8	Motor Volts Error	Drive inhibited	The motor or its wiring is faulty. Check for open or short circuits. Contact your service agent.
9	Other Internal Errors	Drive inhibited	Contact your service agent.
10	Push Too Fast fault	Drive inhibited	The scooter has been pushed faster than the programmed "Push Speed' parameter when the Park Brake Release function has been operated. The scooter has been pushed faster than the programmed "Rollaway Speed" parameter when the Park Brake has been mechanically released. Turn the scooter off and then back on again.

Note:

If you experience any technical problems, it is recommended that you check with your local dealer before attempting to troubleshoot on your own.

The following symptoms could indicate a serious problem with your power scooter. Contact your local dealer if any of the following arises:


- 1. Motor noise
- 2. Frayed harnesses
- 3. Cracked or broken connectors
- 4. Uneven wear on any of tires
- 5. Jerky motion
- 6. Pulling to one side
- 7.Bent or broken wheel assemblies
- 8. Does not power up
- 9. Powers up, but does not move

<u>PG S-Drive controller</u>: PG D-Drive Controller continuously monitors the operating conditions of your scooter. If it detects a problem it will indicate with error message by flashing light. You must count the number of the flash, and see the list to check what kind of error has happened according to the number)

S-Drive Controller – Troubleshooting


1 Bar	The battery needs charging or there is a bad connection to the battery. Check the connections to the battery. If the connections are good, try charging the battery.					
2 Bar	There is a bad connection to the motor. Check all connections between the motor and the controller.					
3 Bar	The motor has a short circuit to a battery connection. Contact your service agent.					
4 Bar	The freewheel switch is activated or the manual brake disengagement mechanism is operated. Check the position of the switch or lever.					
5 Bar	Not used.					
6 Bar	'The S-drive is being inhibited from driving. Inhibit 2 is active'. This may be because the battery charger is connected or the seat is not in the driving position.					
7 Bar	A throttle fault is indicated. Make sure that the throttle is in the rest position before switching on the scooter.					
8 Bar	A controller fault is indicated. Make sure that all connections are secure.					
9 Bar	The parking brakes have a bad connection. Check the parking brake and motor connections. Make sure the controller connections are secure.					
10 Bar	An excessive voltage has been applied to the controller. This is usually caused by a poor battery connection. Check the battery connections.					

CIRCUIT DIAGRAM




HEARTWAY'S PATENT


A brand new double-A arms suspension system.




S12 - BOM LIST DRAWING



S12T - BOM LIST DRAWING



S12X - BOM LIST DRAWING



WARRANTY DECLARATION

Quality/ Warranty Declaration

Products are to be fit for purpose and of excellent quality and performance. For valid warranty claims Heartway will, at their discretion, replace/ repair/ refund items mutually agreed to be defective.

Heartway's Warranty as Following:

- Frame: Two-year limited warranty
- Controllers: One-and-a-half-year limited warranty
- Electronic Components and Charger: One-year limited warranty
- ➢ Warranty Exclusion. The following items are not covered by warranty.

Ŷ	Motor brushes	¢	Wheel Tires	Ŷ	Arm Pads
Ŷ	Seat Cushion	Ŷ	Fuses / Bulbs	Ŷ	Tiller Cover
∻	Rear Shroud	∻	Front Shroud	∻	Batteries and Consumable parts

Any damage or defect of any nature occurring from the misuse, abuse of the product, improper operation or improper storage is not to be covered. The warranty is to start from the date of arrival of our products.



HEARTWAY MEDICAL PRODUCTS CO., LTD.

NO. 6, ROAD 25, TAICHUNG INDUSTRIAL PARK, TAICHUNG. TAIWAN R.O.C.408